

TURBICHEM D-DIMER

(Turbidimetry Method)

KIT NAME	KIT SIZE	CAT. NO
Turbichem - D - Dimer	1 x 40 ml	TDD001040M



INTRODUCTION

D Dimer (D-D) is intended for Invitro quantitative determination of D-D in human plasma. D-dimer is a fibrin degradation product, a small protein fragment present in blood after fibrinolysis degrades a blood clot. D-dimer is normally undetectable in the blood and is synthesized only after a clot has formed and is in the process of being broken down. D-dimer levels rise when a significant formation and breakdown of blood clots occurs. D-dimer is useful in diagnosis of deep vein thrombosis (DVT), pulmonary embolism (PE), and disseminated intravascular coagulation (DIC).

METHOD PRINCIPLE

The Kit utilizes latex-enhanced immunoturbidimetry to measure the D-Dimer level in human plasma. During the test, D-Dimer in the sample binds with the specific anti-d-dimer antibody that is coated on latex particles to cause agglutination. The turbidity caused by agglutination is detected optically by chemistry analyzer. The change in absorbance is proportional to the level of d-dimer in the sample. The actual concentration is obtained by comparing with a calibration curve with known concentrations.

KIT CONTENTS

R1 - D-Dimer Buffer	1 x 30 ml
R2 - D-Dimer antibody	1 x 10 ml
R3 - D-Dimer Calibrator	1 vial

The reagents when stored at 2-8°C are stable up to expiry date printed on the package. The reagents are stable for 10 days on board the analyser at 2-10°C. Protect from light and avoid contamination.

WORKING REAGENT PREPARATION AND STABILITY

Assay can be performed with use of separate R1-D-D and R2-D-D reagents of 3 parts of R1-D-D with 1 part of R2-D-D. Avoid foaming.

CONCENTRATIONS IN THE TEST

R1 - Tris buffer solution. Sodium azide < 0.1%

R2 - Latex suspension, anti-d-dimer antibodies, buffer solution, sodium azide < 0.1%

WARNINGS AND NOTES

1. The Kit is for in vitro diagnostic use only. Not for use in humans or animals.
2. The instructions should be followed to obtain accurate results.
3. Do not use the reagents beyond the expiration date.
4. Treat all specimens as infectious. Proper handling and disposal procedures of specimens and test materials should be strictly followed

ADDITIONAL EQUIPMENT

- Automatic analyzer or photometer able to read at 570 nm
- Thermostat at 37°C
- General laboratory equipment

SPECIMEN

Follow standard laboratory procedures to collect in sodium citrated plasma samples.

It is recommended to perform test immediately after sample collection. If the test cannot be done immediately, store sample at 2-8°C for up to 1 day or at -80°C for up to 6 months. Avoid repeated freezing and thawing.

PLOTTING OF MULTIPOINT CURVE

The Turbichem D-Dimer is based on Non-Linear Reactions, hence it is strongly recommended to run Multi-standard mode to plot the Multi-point curve to have better accuracy and precise result.

Serial Dilution Step

	1st	2nd	3rd	4th	5th
Calibrator	100 µl	50 µl from 1st Tube	50 µl from 2nd Tube	50 µl from 3rd Tube	50 µl from 4th Tube
Normal Saline	0	50 µl	50 µl	50 µl	50 µl
Ratio of Dilution	Neat	1/2	1/4	1/8	1/16

PROCEDURE

These reagents may be used both for manual assay Sample Start and in several automatic analyzers. Applications for them are available on request.

Wavelength	570 nm
Temperature	37°C
Cuvette	1 cm

Pipette into the cuvette:

Reagent	Calibrator (C)	Test(T)
R1 D-Dimer Buffer	750 µl	750 µl
Calibrator	30 µl	
Sample	-	30 µl
Bring up to the temperature of determination. Then add		
R2 - D-Dimer Antibody	250 µl	250 µl

Mix well, after about 10 sec. (37°C) read the absorbance A1 of the test (T) and calibrator (C) against air or water. After exactly 180 secs. (for all temperature) read the absorbance A2 of the test (T) and calibrator (C). Calculate $\Delta A/\text{min}$. ($A2 - A1$) for the test and calibrator.

CALCULATION

D-Dimer concentration = $\Delta A(T) / \Delta A(C) \times \text{calibrator concentration}$

REFERENCE VALUES

upto 0.5 mg/L (0.5 mg/l = 500 ng/ml)

(obtained value in mg/l x 1000 = ng/ml where 1000 is the conversion factor)

It is recommended for each laboratory to establish its own reference ranges for local population.

QUALITY CONTROL

To ensure adequate quality control, each run should include assayed normal and abnormal controls. If commercial controls are not available it is recommended that known value samples be aliquoted, frozen and used as controls

PERFORMANCE CHARACTERISTICS

- **Linearity** : 0 – 20 mg/L (R≥0.990)
- **Precision** : Within Run: CV≤8%; Run-to-Run: Cv≤10%
- **Interference**: no interference detected for: ascorbic acid (50 mg/dL), Bilirubin (19.6 mg/dL), free bilirubin (18.4 mg/dL) Rheumatoid factor(500 IU/ml)and hemoglobin(460 mg/dL).

WASTE MANAGEMENT

Please refer to local legal requirements.

LITERATURE

1. Adam S.S., Key N.S., Greenberg C.S. D-dimer antigen: current concepts and future prospects. Blood 113 (13): 2878-87.
2. Gaffney, P.J. Distinction between Fibrinogen and Fibrin Degradation Products in Plasma. Clin. Chem. Acta. 65 (1): 109-115; 197
3. Rylatt, D.B., et al. An Immunoassay for Human D-Dimer using Monoclonal Antibodies. Thromb. Res. 31(6): 767-778; 1986.
4. Smith, R.T., et al. Fibrin Degradation Products in the Postoperative Period. Evaluation of a New Latex Agglutination Method. Am. J. Clin. Pathol. 60(5): 644-647; 1973

SYSTEM PARAMETERS

Method	Fixed Time (2-Point)
Wavelength	570 nm
Zero Setting	Distilled Water
Temperature Setting	37° C
Incubation Temperature	37° C
Incubation Time	----
Delay Time	10 secs
Read Time	180 secs
No. of Reading	2
Interval Time	----
Sample Volume	0.03 ml (30 ul)
Reagent Volume	1.0 ml (1000 ul)
Standard Concentration	Refer Calibrator Vial
Units	mg/L
Factor	----
Reaction Slope	Increasing
Linearity	20 mg/L



Genuine Biosystem Private Limited
Plot No.97 & 98, kattabomman street,
Parvathy Nagar Extension,
Old Perungalathur, Chennai - 600063, India.
Ph: +91-44-48681845
Email: genuinebiosystem@gmail.com
website: www.genuinebiosystem.com